

Period realizacije: 13– 16. oktobar 2020. godine; 1–5. februar 2021. godine

ZADATAK


Izrada smernica za uspostavljanje **programa osiguranja kvaliteta** u skladu sa zahtevima članova 4, 56, 60, 61, 63, 82, 83 **Direktive Saveta 2013/59/Euratom (BSS)** uzimajući u obzir nacionalnu regulatornu infrastrukturu u oblasti primene izvora zračenja u medicini.

REZULTAT

Sprovedeno je **istraživanje** o frekvencijama kontrole kvaliteta (QC) ispitivanjem određenih parametara za izvore zračenja koji se koriste u medicini.

Analiza rezultata istraživanja frekvencija kontrole kvaliteta određenih parametara za izvore zračenja koji se koriste u nuklearnoj medicini, radioterapiji i za rendgen aparate za kompjuterizovanu tomografiju.

Smernice za uspostavljanje programa osiguranja kvaliteta u formi **Pravilnika za uspostavljanje programa osiguranja kvaliteta**.



ZAKLJUČCI & PREPORUKE

Komentari primljeni od anketiranih institucija sadrže opšti zahtev za većim brojem medicinskih fizičara koji sprovode redovne kontrole kvaliteta. Takođe je primećeno da institucijama nedostaje oprema za vršenje različitih internih provera kvaliteta.

Preporučuje se da SRBATOM blagovremeno primeni **Pravilnik za uspostavljanje programa osiguranja kvaliteta**. Njegova primena će vremenom dovesti do poboljšanja nedostataka u ovoj oblasti.

Official Journal
of the European Union

L 13
17 January 2014

Legislation

Volume 37

Contents

II: Non-legislative acts

DIRECTIVES

* Council Directive 2013/59/Euratom of 5 December 2013 laying down basic safety standards for protection against the dangers arising from exposure to ionizing radiation, and repealing Directives 89/686/Euratom, 90/641/Euratom, 92/71/Euratom, 97/41/Euratom and 2000/13/Euratom